



### Administrative Data

Institution SZX - Institut für Radioonkologie  
Physicist Krauss  
PatientID zz-Inhomogenitätsphantom  
Patient Name zz-Inhomogenitätsphantom, zz-PTW-T40037, zz-Feb-2013  
Comment k(0PTPN6X20AXB10W)=0.999

### Data Set A

Q:\TrueBeam\\_COMMISSIONING\VeriSoft\2013-02-05 T40037\0BPB6X20AXB10W.dcm

### Data Set B

Q:\TrueBeam\\_COMMISSIONING\VeriSoft\2013-02-05 T40037\0BPB6X20b.mcc

### Manipulations

Command	Data Set	Parameters	Value
Calibrate Air Density	B	kUser	0.988
Calibrate Air Density	B	kUser	0.999

### Gamma 2D - Parameters

3.0 mm Distance- To- Agreement  
3.0 % Dose Difference with ref. to Max. dose of measured data set

### Statistics

Number of Dose Points	729
Evaluated Dose Points	729 ( 100.0 %)
Passed	727 ( 99.7 %)
Failed	2 ( 0.3 %)
Result	99.7 % (Green)

### Settings

Passing criteria	Gamma $\leq$ 1.0
Green	90.0 % to 100.0 %
Yellow	75.0 % to 90.0 %
Red	0.0 % to 75.0 %

