



### Administrative Data

Institution SZX - Institut für Radioonkologie  
Physicist Krauss  
PatientID zz-Inhomogenitätsphantom  
Patient Name zz-Inhomogenitätsphantom, zz-PTW-T40037, zz-Feb-2013  
Comment k(0PTPN6X20AAA)=0.988

### Data Set A

Q:\TrueBeam\\_COMMISSIONING\VeriSoft\2013-02-05 T40037\0PLP6X20AAA.dcm

### Data Set B

Q:\TrueBeam\\_COMMISSIONING\VeriSoft\2013-02-05 T40037\0PLP6X20a.mcc

### Manipulations

| Command               | Data Set | Parameters | Value |
|-----------------------|----------|------------|-------|
| Calibrate Air Density | B        | kUser      | 0.988 |

### Gamma 2D - Parameters

2.0 mm Distance- To- Agreement  
2.0 % Dose Difference with ref. to Max. dose of measured data set

### Statistics

|                       |                 |
|-----------------------|-----------------|
| Number of Dose Points | 729             |
| Evaluated Dose Points | 729 ( 100.0 %)  |
| Passed                | 578 ( 79.3 %)   |
| Failed                | 151 ( 20.7 %)   |
| Result                | 79.3 % (Yellow) |

### Settings

|                  |                   |
|------------------|-------------------|
| Passing criteria | Gamma $\leq$ 1.0  |
| Green            | 90.0 % to 100.0 % |
| Yellow           | 75.0 % to 90.0 %  |
| Red              | 0.0 % to 75.0 %   |

