



### Administrative Data

Institution SZX - Institut für Radioonkologie  
Physicist Krauss  
PatientID zz-Inhomogenitätsphantom  
Patient Name zz-Inhomogenitätsphantom, zz-PTW-T40037, zz-Feb-2013  
Comment Calc grid 1.5mm, k(0PTPN6X20AAA)=0.988

### Data Set A

Q:\TrueBeam\\_COMMISSIONING\VeriSoft\2013-02-05 T40037\1P0P6X20AAAh.dcm

### Data Set B

Q:\TrueBeam\\_COMMISSIONING\VeriSoft\2013-02-05 T40037\1P0P6X20b.mcc

### Manipulations

Command	Data Set	Parameters	Value
Calibrate Air Density	B	kUser	0.988

### Gamma 2D - Parameters

3.0 mm Distance- To- Agreement  
3.0 % Dose Difference with ref. to Max. dose of measured data set

### Statistics

Number of Dose Points	729
Evaluated Dose Points	729 ( 100.0 %)
Passed	701 ( 96.2 %)
Failed	28 ( 3.8 %)
Result	96.2 % (Green)

### Settings

Passing criteria	Gamma $\leq 1.0$
Green	90.0 % to 100.0 %
Yellow	75.0 % to 90.0 %
Red	0.0 % to 75.0 %

