



### Administrative Data

Institution SZX - Institut für Radioonkologie  
Physicist Krauss  
PatientID zz-Inhomogenitätsphantom  
Patient Name zz-Inhomogenitätsphantom, zz-PTW-T40037, zz-Feb-2013  
Comment k(0PTPN6X20AXB10M)=0.988

### Data Set A

Q:\TrueBeam\\_COMMISSIONING\VeriSoft\2013-02-05 T40037\1P0P6X20AXB10M.dcm

### Data Set B

Q:\TrueBeam\\_COMMISSIONING\VeriSoft\2013-02-05 T40037\1P0P6X20b.mcc

### Manipulations

Command	Data Set	Parameters	Value
Calibrate Air Density	B	Pressure	976.00 hPa
		Temperature	23.2 °C
Calibrate Air Density	B	kUser	0.988

### Gamma 2D - Parameters

2.0 mm Distance- To- Agreement  
2.0 % Dose Difference with ref. to Max. dose of measured data set

### Statistics

Number of Dose Points 729  
Evaluated Dose Points 729 ( 100.0 %)  
Passed 657 ( 90.1 %)  
Failed 72 ( 9.9 %)  
Result 90.1 % (Green)

### Settings

Passing criteria Gamma  $\leq 1.0$   
Green 90.0 % to 100.0 %  
Yellow 75.0 % to 90.0 %  
Red 0.0 % to 75.0 %

