



Administrative Data

Institution SZX - Institut für Radioonkologie
Physicist Krauss
PatientID zz-Inhomogenitätsphantom
Patient Name zz-Inhomogenitätsphantom, zz-PTW-T40037, zz-Feb-2013
Comment Calc grid 1.5mm, k(0PTPN6X20AXB10W)=0.999

Data Set A

Q:\TrueBeam_COMMISSIONING\VeriSoft\2013-02-05 T40037\1P0P6X20AXB10Wh.dcm

Data Set B

Q:\TrueBeam_COMMISSIONING\VeriSoft\2013-02-05 T40037\1P0P6X20b.mcc

Manipulations

Command	Data Set	Parameters	Value
Calibrate Air Density	B	Pressure	976.00 hPa
		Temperature	23.2 °C
Calibrate Air Density	B	kUser	0.988
Calibrate Air Density	B	kUser	0.999

Gamma 2D - Parameters

2.0 mm Distance- To- Agreement
2.0 % Dose Difference with ref. to Max. dose of measured data set

Statistics

Number of Dose Points	729
Evaluated Dose Points	729 (100.0 %)
Passed	692 (94.9 %)
Failed	37 (5.1 %)
Result	94.9 % (Green)

Settings

Passing criteria	Gamma \leq 1.0
Green	90.0 % to 100.0 %
Yellow	75.0 % to 90.0 %
Red	0.0 % to 75.0 %

