



Administrative Data

Institution SZX - Institut für Radioonkologie
Physicist Krauss
PatientID zz-Inhomogenitätsphantom
Patient Name zz-Inhomogenitätsphantom, zz-PTW-T40037, zz-Feb-2013
Comment k(0PTPN6X20AAA)=0.988

Data Set A

Q:\TrueBeam_COMMISSIONING\VeriSoft\2013-02-05 T40037\1TBLN6X20AAA.dcm

Data Set B

Q:\TrueBeam_COMMISSIONING\VeriSoft\2013-02-05 T40037\1TBL6X20b.mcc

Manipulations

Command	Data Set	Parameters	Value
Calibrate Air Density	B	kUser	0.988

Gamma 2D - Parameters

2.0 mm Distance- To- Agreement
2.0 % Dose Difference with ref. to Max. dose of measured data set

Statistics

Number of Dose Points	729
Evaluated Dose Points	729 (100.0 %)
Passed	565 (77.5 %)
Failed	164 (22.5 %)
Result	77.5 % (Yellow)

Settings

Passing criteria	Gamma \leq 1.0
Green	90.0 % to 100.0 %
Yellow	75.0 % to 90.0 %
Red	0.0 % to 75.0 %

