



Administrative Data

Institution SZX - Institut für Radioonkologie
Physicist Krauss
PatientID zz-Inhomogenitätsphantom
Patient Name zz-Inhomogenitätsphantom, zz-PTW-T40037, zz-Feb-2013
Comment Calc grid 1.5mm, k(0PTPN6X20AAA)=0.988

Data Set A

Q:\TrueBeam_COMMISSIONING\VeriSoft\2013-02-05 T40037\1TBLN6X20AAAh.dcm

Data Set B

Q:\TrueBeam_COMMISSIONING\VeriSoft\2013-02-05 T40037\1TBL6X20b.mcc

Manipulations

Command	Data Set	Parameters	Value
Calibrate Air Density	B	kUser	0.999
Calibrate Air Density	B	kUser	0.988
Calibrate Air Density	B		

Gamma 2D - Parameters

3.0 mm Distance- To- Agreement
3.0 % Dose Difference with ref. to Max. dose of measured data set

Statistics

Number of Dose Points 729
Evaluated Dose Points 729 (100.0 %)
Passed 697 (95.6 %)
Failed 32 (4.4 %)
Result 95.6 % (Green)

Settings

Passing criteria Gamma \leq 1.0
Green 90.0 % to 100.0 %
Yellow 75.0 % to 90.0 %
Red 0.0 % to 75.0 %

