

SQUARE FIELD OUTPUT FACTORS: 60008-11

CLINAC: 2300C/D
S/N: 156

ENERGY: 15MV
DATE: 12 Feb. 2009

PHANTOM: MP3 with Trifix
VOLTAGE: 0V

SSD: 90 cm
DEPTH: 10 cm

RAW DATA

MEASUREMENT	1X1	2X2	3X3	4X4	5X5	7X7	10X10	15X15	20X20	30X30	40X40
1	0.6062	0.7395	0.7929	0.8244	0.8463	0.8795	0.9121	0.9459	0.9703	1.0070	1.0310
2	0.6062	0.7397	0.7926	0.8252	0.8465	0.8791	0.9123	0.9467	0.9703	1.0060	1.0310
3	0.6058	0.7386	0.7920	0.8250	0.8457	0.8784	0.9116	0.9460	0.9696	1.0060	1.0310
4	0.6063	0.7388	0.7919	0.8241	0.8461	0.8788	0.9116	0.9460	0.9700		
5			0.7914	0.8241			0.9115	0.9458			
6							0.9111				
7							0.9117				
8							0.9117				
9											
10											
11											
12											
13											
Average	0.6061	0.7392	0.7922	0.8246	0.8462	0.8790	0.9117	0.9461	0.9701	1.0063	1.0310
Std Dev	0.0002	0.0005	0.0006	0.0005	0.0003	0.0005	0.0004	0.0004	0.0003	0.0006	0.0000
SDOM	0.0001	0.0003	0.0003	0.0002	0.0002	0.0002	0.0001	0.0002	0.0002	0.0003	0.0000

OUTPUT FACTORS

FIELD	1	2	3	4	5	7	10	15	20	30	40
OF	0.6648	0.8107	0.8689	0.9044	0.9281	0.9641	1.0000	1.0377	1.0640	1.1038	1.1309
SDOM	0.0002	0.0003	0.0003	0.0003	0.0002	0.0003	0.0002	0.0002	0.0002	0.0004	0.0001
SDOM [%]	0.03	0.04	0.03	0.03	0.02	0.03	0.02	0.02	0.02	0.03	0.01

OF 15MV 60008-11

