

SQUARE FIELD OUTPUT FACTORS: 60012-206

CLINAC: 2300C/D
S/N: 156

ENERGY: 15MV
DATE: 12 Feb. 2009

PHANTOM: MP3 with Trifix
VOLTAGE: 0V

SSD: 90 cm
DEPTH: 10 cm

RAW DATA

MEASUREMENT	1X1	2X2	3X3	4X4	5X5	7X7	10X10	15X15	20X20	30X30	40X40
1	0.6312	0.7842	0.8478	0.8844	0.9108	0.9482	0.9891	1.0310	1.0640	1.1160	1.1530
2	0.6313	0.7849	0.8476	0.8844	0.9107	0.9482	0.9891	1.0300	1.0640	1.1160	1.1520
3	0.6308	0.7814	0.8440	0.8816	0.9078	0.9458	0.9893	1.0300	1.0620	1.1150	1.1520
4	0.6307	0.7814	0.8443	0.8813	0.9084	0.9460	0.9870	1.0300	1.0620	1.1140	
5	0.6301	0.7822	0.8444		0.9083		0.9873				
6	0.6303	0.7819	0.8444				0.9855				
7							0.9851				
8							0.9853				
9							0.9849				
10							0.9852				
11							0.9846				
12											
13											
Average	0.6307	0.7827	0.8454	0.8829	0.9092	0.9471	0.9866	1.0303	1.0630	1.1153	1.1523
Std Dev	0.0005	0.0015	0.0018	0.0017	0.0014	0.0013	0.0019	0.0005	0.0012	0.0010	0.0006
SDOM	0.0002	0.0006	0.0007	0.0009	0.0006	0.0007	0.0006	0.0002	0.0006	0.0005	0.0003

OUTPUT FACTORS

FIELD	1	2	3	4	5	7	10	15	20	30	40
OF	0.6393	0.7933	0.8569	0.8949	0.9216	0.9599	1.0000	1.0443	1.0775	1.1304	1.1680
SDOM	0.0006	0.0008	0.0009	0.0010	0.0009	0.0009	0.0008	0.0006	0.0008	0.0007	0.0007
SDOM [%]	0.09	0.10	0.11	0.11	0.09	0.09	0.08	0.06	0.07	0.07	0.06

OF 15MV 60012-206

