

SQUARE FIELD OUTPUT FACTORS

SITE: KFJ Vienna
ENERGY: 15MV

CLINAC: 2300C/D
PHANTOM: MP3 with Trifix

S/N: 156
SSD: 90 cm

DATE: 12 Feb. 2009
DEPTH: 10 cm

OUTPUT FACTORS

FIELD	1	2	3	4	5	7	10	15	20	30	40
31002	0.557	0.776	0.859	0.901	0.928	0.965	1.000	1.037	1.062	1.099	1.125
60012	0.639	0.793	0.857	0.895	0.922	0.960	1.000	1.044	1.077	1.130	1.168
60008	0.665	0.811	0.869	0.904	0.928	0.964	1.000	1.038	1.064	1.104	1.131
31016	0.596	0.788	0.863	0.903	0.928	0.965	1.000	1.038	1.064	1.107	1.150
31003	0.383	0.736	0.846	0.897	0.926	0.965	1.000	1.037	1.062	1.098	1.122
31018	0.639	0.804	0.870	0.908	0.932	0.967	1.000	1.033	1.055	1.086	1.105

