

SQUARE FIELD OUTPUT FACTORS

SITE: KFJ Vienna
ENERGY: 6MV

CLINAC: 2300C/D
PHANTOM: MP3 with Trifix

S/N: 156
SSD: 90 cm

DATE: 12 Feb. 2009
DEPTH: 10 cm

OUTPUT FACTORS

FIELD	1	2	3	4	5	7	10	15	20	30	40
31002	0.624	0.783	0.830	0.865	0.894	0.943	1.000	1.059	1.101	1.157	1.194
60012	0.698	0.781	0.820	0.854	0.884	0.937	1.000	1.074	1.129	1.214	1.273
60008	0.725	0.799	0.835	0.868	0.897	0.945	1.000	1.061	1.104	1.164	1.204
31016	0.660	0.786	0.832	0.865	0.894	0.944	1.000	1.063	1.107	1.174	1.232
31003	0.410	0.752	0.826	0.864	0.893	0.944	1.000	1.060	1.100	1.154	1.189
31018	0.711	0.804	0.843	0.876	0.904	0.950	1.000	1.053	1.086	1.129	1.155

