

QUOTIENT OF OUTPUT FACTORS FOR SQUARE FIELDS

SITE: KFJ Vienna
ENERGY: 15MV

CLINAC: 2300C/D
PHANTOM: MP3 with Trifix

S/N: 156
SSD: 90 cm

DATE: 12 Feb. 2009
DEPTH: 10 cm

OUTPUT FACTOR QUOTIENT 15MV: OF(DETECTOR) / OF(31002-368)

FIELD	1	2	3	4	5	7	10	15	20	30	40
31002	1,000	1,000	1,000	1,000	1,000	1,000	1,000	1,000	1,000	1,000	1,000
60012	1,149	1,022	0,998	0,993	0,993	0,995	1,000	1,007	1,014	1,029	1,038
60008	1,194	1,045	1,012	1,004	1,000	0,999	1,000	1,001	1,002	1,004	1,005
31016	1,071	1,016	1,005	1,002	1,001	1,000	1,000	1,001	1,002	1,007	1,022
31003	0,688	0,948	0,985	0,995	0,999	1,000	1,000	1,000	0,999	0,999	0,997
31018	1,148	1,036	1,013	1,007	1,005	1,002	1,000	0,996	0,994	0,988	0,982

OUTPUT FACTOR QUOTIENT: OF(DETECTOR) / OF(31002-368)

