

# QUOTIENT OF OUTPUT FACTORS FOR SQUARE FIELDS

**SITE:** KFJ Vienna  
**ENERGY:** 6MV

**CLINAC:** 2300C/D  
**PHANTOM:** MP3 with Trifix

**S/N:** 156  
**SSD:** 90 cm

**DATE:** 12 Feb. 2009  
**DEPTH:** 10 cm

**OUTPUT FACTOR QUOTIENT 6MV: OF(DETECTOR) / OF(31002-368)**

FIELD	1	2	3	4	5	7	10	15	20	30	40
31002	1,000	1,000	1,000	1,000	1,000	1,000	1,000	1,000	1,000	1,000	1,000
60012	1,118	0,997	0,987	0,988	0,989	0,993	1,000	1,014	1,025	1,049	1,066
60008	1,161	1,020	1,006	1,004	1,003	1,001	1,000	1,002	1,002	1,006	1,009
31016	1,057	1,004	1,002	1,000	0,999	1,001	1,000	1,004	1,005	1,014	1,032
31003	0,657	0,960	0,995	0,999	0,999	1,000	1,000	1,001	0,999	0,997	0,996
31018	1,139	1,027	1,015	1,012	1,011	1,006	1,000	0,995	0,986	0,976	0,967

