

Wedge factor report

Institute:	KFJ	Department:	IRO
Patient ID:	zz-Halbjahrescheck	ID QA:	Halbjahrescheck5132X10
Plan:	5132X10	LINAC:	TrueBeam_BLAU 10 MV
Detector:	aS1000	Acquisition date:	2015/01/14
Epiqa calib date :	2014/06/26	Epiqa calib depth :	mix
Description:	1.Halbjahrescheck 2015 Blau		

Results

	X size [cm]	Y size [cm]	Wedge type	Wedge angle	Wedge orient	Gy/100MU TPS	Gy/100MU PV	Diff %	WF TPS	WF PV	Diff %
15IN	10.00	10.00	DYNAMIC	15.00	IN	0.980	0.985	-0.5	0.937	0.938	-0.2
45OUT	10.00	10.00	DYNAMIC	45.00	OUT	0.839	0.843	-0.4	0.802	0.802	-0.1
OPEN10	10.00	10.00	NONE	#####	####	1.046	1.050	-0.4	REF ECL	REF PV	

Notes: _____

Approval: _____

Physicist: _____